


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD122420\  
Data File : VD067942.D  
Acq On : 24 Dec 2020 13:14  
Operator : VA/SY  
Sample : BFB  
Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_D  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D122220S.M  
Title : Sw846 8260  
Last Update : wed Dec 23 07:34:16 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1811

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw    | Result    |
|--------|---------|--------|--------|-------|--------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn    | Pass/Fail |
| 50     | 95      | 15     | 40     | 18.6  | 50787  | PASS      |
| 75     | 95      | 30     | 60     | 51.2  | 139931 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 273088 | PASS      |
| 96     | 95      | 5      | 9      | 6.4   | 17408  | PASS      |
| 173    | 174     | 0.00   | 2      | 1.9   | 3957   | PASS      |
| 174    | 95      | 50     | 100    | 76.2  | 208192 | PASS      |
| 175    | 174     | 5      | 9      | 7.7   | 15957  | PASS      |
| 176    | 174     | 95     | 101    | 96.3  | 200448 | PASS      |
| 177    | 176     | 5      | 9      | 6.8   | 13574  | PASS      |