



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD010923\
 Data File : VD075086.D
 Acq On : 09 Jan 2023 10:41
 Operator : KP/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Title : SW846 8260
 Last Update : Tue Jan 10 03:36:33 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	12669	PASS
75	95	30	60	52.6	32211	PASS
95	95	100	100	100.0	61248	PASS
96	95	5	9	6.4	3927	PASS
173	174	0.00	2	1.1	563	PASS
174	95	50	100	82.0	50219	PASS
175	174	5	9	7.6	3792	PASS
176	174	95	101	96.5	48469	PASS
177	176	5	9	6.9	3336	PASS