

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013023\
 Data File : VD075236.D
 Acq On : 30 Jan 2023 14:24
 Operator : KP/SY
 Sample : 01348-02
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 E-H5(31.5-33.5)

Quant Time: Jan 31 00:22:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	78783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	143485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	137935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	62330	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	40086	49.863	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.720%
35) Dibromofluoromethane	7.804	113	43393	50.048	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.100%
50) Toluene-d8	10.269	98	178913	52.408	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.820%
62) 4-Bromofluorobenzene	12.575	95	56295	51.959	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013023\
Data File : VD075236.D
Acq On : 30 Jan 2023 14:24
Operator : KP/SY
Sample : 01348-02
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
E-H5(31.5-33.5)

Quant Time: Jan 31 00:22:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

