

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013023\
 Data File : VD075240.D
 Acq On : 30 Jan 2023 16:15
 Operator : KP/SY
 Sample : 01348-06
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-H5(30.5-32)

Quant Time: Jan 31 00:22:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	52852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	98920	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	100513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	46872	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	32347	59.978	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.960%
35) Dibromofluoromethane	7.810	113	31815	53.225	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.460%
50) Toluene-d8	10.269	98	124814	53.032	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.060%
62) 4-Bromofluorobenzene	12.575	95	42019	56.255	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 31 00:22:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

