

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013023\
 Data File : VD075243.D
 Acq On : 30 Jan 2023 17:39
 Operator : KP/SY
 Sample : 01360-01
 Misc : 5.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3-WC

Quant Time: Jan 31 00:23:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 19 01:32:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	57643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	107404	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	108445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	50270	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	34953	59.423	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.840%
35) Dibromofluoromethane	7.804	113	34386	52.982	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.960%
50) Toluene-d8	10.269	98	134273	52.545	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.080%
62) 4-Bromofluorobenzene	12.575	95	45679	56.324	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013023\
Data File : VD075243.D
Acq On : 30 Jan 2023 17:39
Operator : KP/SY
Sample : 01360-01
Misc : 5.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3-WC

Quant Time: Jan 31 00:23:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 19 01:32:27 2023
Response via : Initial Calibration

