

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022223\
 Data File : VD075401.D
 Acq On : 22 Feb 2023 16:03
 Operator : KP/SY
 Sample : 01639-01
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-02212023

Quant Time: Feb 22 22:38:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 21 16:26:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	1470	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	1457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	577	88.703	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 177.400%#	
35) Dibromofluoromethane	7.804	113	527	60.390	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 120.780%	
50) Toluene-d8	10.263	98	1706	54.170	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.340%	
62) 4-Bromofluorobenzene	12.586	95	555	52.482	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.960%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022223\
Data File : VD075401.D
Acq On : 22 Feb 2023 16:03
Operator : KP/SY
Sample : 01639-01
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-02212023

Quant Time: Feb 22 22:38:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 21 16:26:04 2023
Response via : Initial Calibration

