

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022223\
 Data File : VD075402.D
 Acq On : 22 Feb 2023 16:31
 Operator : KP/SY
 Sample : 01636-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-022123

Quant Time: Feb 22 22:39:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 21 16:26:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.863	168	256	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.781	114	879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	719	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	113	50.000	ug/l	#-0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	434	206.855	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	413.700%#
35) Dibromofluoromethane	7.799	113	341	65.349	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	130.700%
50) Toluene-d8	10.263	98	1172	62.235	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	124.480%
62) 4-Bromofluorobenzene	12.569	95	262	41.433	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022223\
Data File : VD075402.D
Acq On : 22 Feb 2023 16:31
Operator : KP/SY
Sample : 01636-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-022123

Quant Time: Feb 22 22:39:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 21 16:26:04 2023
Response via : Initial Calibration

