

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
 Data File : VD072238.D
 Acq On : 17 Mar 2022 11:56
 Operator : VA/SY
 Sample : VD0317SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0317SBL01

Quant Time: Mar 17 17:58:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	482490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	779907	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	813301	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	405646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	219739	40.866	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.740%
35) Dibromofluoromethane	7.914	113	236638	43.117	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.240%
50) Toluene-d8	10.338	98	952553	47.466	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.940%
62) 4-Bromofluorobenzene	12.626	95	392008	50.274	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.540%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
Data File : VD072238.D
Acq On : 17 Mar 2022 11:56
Operator : VA/SY
Sample : VD0317SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0317SBL01

Quant Time: Mar 17 17:58:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

