

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
 Data File : VD072242.D
 Acq On : 17 Mar 2022 13:49
 Operator : VA/SY
 Sample : N1922-01
 Misc : 5.59G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 7PH-1-7.5-8

Quant Time: Mar 18 04:29:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	452823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	752840	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	801041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	406685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	230620	45.699	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.400%
35) Dibromofluoromethane	7.908	113	232183	43.826	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.660%
50) Toluene-d8	10.338	98	926076	47.806	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.620%
62) 4-Bromofluorobenzene	12.626	95	414950	55.129	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.260%
Target Compounds						
16) Acetone	4.150	43	11668	10.649	ug/l	99
17) Carbon Disulfide	4.409	76	6963	2.196	ug/l #	82
95) Naphthalene	15.384	128	17713	1.171	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
Data File : VD072242.D
Acq On : 17 Mar 2022 13:49
Operator : VA/SY
Sample : N1922-01
Misc : 5.59G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
7PH-1-7.5-8

Quant Time: Mar 18 04:29:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

