

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
 Data File : VD072245.D
 Acq On : 17 Mar 2022 15:14
 Operator : VA/SY
 Sample : N1913-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BS-SOILP

Quant Time: Mar 18 04:30:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	279739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	465156	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	514858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	266130	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	168064	53.909	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.820%	
35) Dibromofluoromethane	7.914	113	154646	47.244	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 94.480%	
50) Toluene-d8	10.338	98	577198	48.224	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 96.440%	
62) 4-Bromofluorobenzene	12.626	95	256110	55.070	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.140%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
Data File : VD072245.D
Acq On : 17 Mar 2022 15:14
Operator : VA/SY
Sample : N1913-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BS-SOILP

Quant Time: Mar 18 04:30:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

