

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
 Data File : VD072246.D
 Acq On : 17 Mar 2022 15:42
 Operator : VA/SY
 Sample : N2007-01
 Misc : 5.59G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP1

Quant Time: Mar 18 04:31:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	424037	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	712766	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	726241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	314566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	225109	47.636	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.280%
35) Dibromofluoromethane	7.909	113	227803	45.417	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.840%
50) Toluene-d8	10.338	98	869789	47.425	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.860%
62) 4-Bromofluorobenzene	12.626	95	363851	51.058	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.120%
Target Compounds						
17) Carbon Disulfide	4.415	76	7403	2.291	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
Data File : VD072246.D
Acq On : 17 Mar 2022 15:42
Operator : VA/SY
Sample : N2007-01
Misc : 5.59G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP1

Quant Time: Mar 18 04:31:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

