

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
 Data File : VD072261.D
 Acq On : 18 Mar 2022 00:34
 Operator : VA/SY
 Sample : VD0317SBL02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0317SBL02

Quant Time: Mar 18 05:38:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	380628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	643319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	688936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	340126	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	191039	45.036	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.080%
35) Dibromofluoromethane	7.908	113	203255	44.898	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.800%
50) Toluene-d8	10.338	98	806022	48.692	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.380%
62) 4-Bromofluorobenzene	12.620	95	329536	51.235	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.460%
Target Compounds						
17) Carbon Disulfide	4.415	76	5858	2.197	ug/l	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
Data File : VD072261.D
Acq On : 18 Mar 2022 00:34
Operator : VA/SY
Sample : VD0317SBL02
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0317SBL02

Quant Time: Mar 18 05:38:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

