

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
 Data File : VD072279.D
 Acq On : 18 Mar 2022 09:04
 Operator : VA/SY
 Sample : N1916-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CONCRETE-COMP

Quant Time: Mar 18 09:43:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	368037	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	619835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	667886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	329268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	180393	43.982	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.960%
35) Dibromofluoromethane	7.914	113	122093	27.991	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	55.980%
50) Toluene-d8	10.338	98	784026	49.158	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.626	95	315293	50.878	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.760%
Target Compounds						
16) Acetone	4.156	43	12578	14.124	ug/l	90
17) Carbon Disulfide	4.414	76	3907	1.981	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031722\
Data File : VD072279.D
Acq On : 18 Mar 2022 09:04
Operator : VA/SY
Sample : N1916-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CONCRETE-COMP

Quant Time: Mar 18 09:43:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

