

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
 Data File : VD075626.D
 Acq On : 11 Apr 2023 18:08
 Operator : KP/SY
 Sample : 02260-03
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1A

Quant Time: Apr 12 04:00:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 18:54:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	4072	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	8395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	10194	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.522	152	4156	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	3645	109.116	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	218.240%#
35) Dibromofluoromethane	7.804	113	4527	92.013	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	184.020%#
50) Toluene-d8	10.275	98	10449	54.146	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.300%
62) 4-Bromofluorobenzene	12.575	95	3746	62.934	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
Data File : VD075626.D
Acq On : 11 Apr 2023 18:08
Operator : KP/SY
Sample : 02260-03
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1A

Quant Time: Apr 12 04:00:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 18:54:55 2023
Response via : Initial Calibration

