

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
 Data File : VD075629.D
 Acq On : 11 Apr 2023 19:38
 Operator : KP/SY
 Sample : 02260-12
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1D

Quant Time: Apr 12 02:59:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 18:54:55 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	14678	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	37604	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	47673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	22099	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	13556	112.581	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	225.160%	#
35) Dibromofluoromethane	7.805	113	17193	78.015	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	156.020%	#
50) Toluene-d8	10.269	98	62758	72.602	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	145.200%	#
62) 4-Bromofluorobenzene	12.575	95	18033	67.761	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	135.520%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.799	84	3561	12.115	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
Data File : VD075629.D
Acq On : 11 Apr 2023 19:38
Operator : KP/SY
Sample : 02260-12
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1D

Quant Time: Apr 12 02:59:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 18:54:55 2023
Response via : Initial Calibration

