

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
 Data File : VD075631.D
 Acq On : 11 Apr 2023 20:37
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 12 03:00:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 18:54:55 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.353	0.318	9.9	78	0.00
3 P	Chloromethane	0.614	0.614	0.0	87	0.00
4 C	Vinyl Chloride	0.860	0.827	3.8#	80	0.00
5 T	Bromomethane	0.685	0.614	10.4	76	0.00
6 T	Chloroethane	0.603	0.606	-0.5	86	0.00
7 T	Trichlorofluoromethane	0.734	0.707	3.7	83	-0.01
8 T	Diethyl Ether	0.227	0.238	-4.8	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.476	1.4	85	0.00
10 T	Methyl Iodide	0.577	0.572	0.9	81	0.00
11 T	Tert butyl alcohol	0.019	0.020	-5.3	95	0.00
12 CM	1,1-Dichloroethene	0.463	0.449	3.0#	83	0.00
13 T	Acrolein	0.030	0.028	6.7	90	0.00
14 T	Allyl chloride	0.418	0.418	0.0	86	0.00
15 T	Acrylonitrile	0.087	0.096	-10.3	93	0.00
16 T	Acetone	0.075	0.082	-9.3	94	0.00
17 T	Carbon Disulfide	1.200	1.102	8.2	78	-0.01
18 T	Methyl Acetate	0.226	0.247	-9.3	94	-0.01
19 T	Methyl tert-butyl Ether	0.980	1.057	-7.9	90	-0.01
20 T	Methylene Chloride	0.747	0.676	9.5	86	0.00
21 T	trans-1,2-Dichloroethene	0.547	0.544	0.5	84	0.00
22 T	Diisopropyl ether	1.026	1.147	-11.8	92	0.00
23 T	Vinyl Acetate	0.498	0.553	-11.0	90	0.00
24 P	1,1-Dichloroethane	0.847	0.889	-5.0	89	0.00
25 T	2-Butanone	0.097	0.108	-11.3	96	0.00
26 T	2,2-Dichloropropane	0.777	0.768	1.2	84	0.00
27 T	cis-1,2-Dichloroethene	0.644	0.664	-3.1	87	0.00
28 T	Bromochloromethane	0.290	0.318	-9.7	95	0.00
29 T	Tetrahydrofuran	0.051	0.060	-17.6	96	0.00
30 C	Chloroform	0.964	1.035	-7.4#	90	0.00
31 T	Cyclohexane	0.597	0.553	7.4	82	0.00
32 T	1,1,1-Trichloroethane	0.802	0.820	-2.2	85	0.00
33 S	1,2-Dichloroethane-d4	0.410	0.457	-11.5	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.293	0.323	-10.2	93	0.00
36 T	1,1-Dichloropropene	0.374	0.384	-2.7	83	0.00
37 T	Ethyl Acetate	0.113	0.130	-15.0	97	0.00
38 T	Carbon Tetrachloride	0.359	0.371	-3.3	85	0.00
39 T	Methylcyclohexane	0.446	0.436	2.2	82	0.00
40 TM	Benzene	1.206	1.300	-7.8	88	0.00
41 T	Methacrylonitrile	0.063	0.070	-11.1	96	0.00
42 TM	1,2-Dichloroethane	0.280	0.311	-11.1	92	0.00
43 T	Isopropyl Acetate	0.215	0.245	-14.0	95	0.00
44 TM	Trichloroethene	0.353	0.347	1.7	85	0.00
45 C	1,2-Dichloropropane	0.281	0.302	-7.5#	91	0.00
46 T	Dibromomethane	0.175	0.192	-9.7	91	0.00
47 T	Bromodichloromethane	0.422	0.453	-7.3	91	0.00
48 T	Methyl methacrylate	0.097	0.101	-4.1	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
 Data File : VD075631.D
 Acq On : 11 Apr 2023 20:37
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 12 03:00:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 18:54:55 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	0.00
50 S	Toluene-d8	1.149	1.243	-8.2	91	0.00
51 T	4-Methyl-2-Pentanone	0.116	0.135	-16.4	98	0.00
52 CM	Toluene	0.814	0.836	-2.7#	85	0.00
53 T	t-1,3-Dichloropropene	0.383	0.414	-8.1	90	0.00
54 T	cis-1,3-Dichloropropene	0.466	0.493	-5.8	89	0.00
55 T	1,1,2-Trichloroethane	0.258	0.288	-11.6	95	0.00
56 T	Ethyl methacrylate	0.248	0.282	-13.7	93	0.00
57 T	1,3-Dichloropropane	0.388	0.422	-8.8	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.120	0.142	-18.3	102	0.00
59 T	2-Hexanone	0.081	0.093	-14.8	95	0.00
60 T	Dibromochloromethane	0.308	0.327	-6.2	91	0.00
61 T	1,2-Dibromoethane	0.246	0.258	-4.9	82	0.00
62 S	4-Bromofluorobenzene	0.338	0.382	-13.0	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.301	0.306	-1.7	87	0.00
65 PM	Chlorobenzene	0.971	0.991	-2.1	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.343	0.351	-2.3	77	0.00
67 C	Ethyl Benzene	1.601	1.657	-3.5#	78	0.00
68 T	m/p-Xylenes	0.660	0.693	-5.0	79	0.00
69 T	o-Xylene	0.594	0.637	-7.2	75	0.00
70 T	Styrene	1.018	1.161	-14.0	79	0.00
71 P	Bromoform	0.177	0.196	-10.7	76	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.103	3.034	2.2	76	0.00
74 T	N-ethyl acetate	0.435	0.457	-5.1	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.580	0.614	-5.9	86	0.00
76 T	1,2,3-Trichloropropane	0.325	0.415	-27.7#	111	0.00
77 T	Bromobenzene	0.783	0.767	2.0	74	0.00
78 T	n-propylbenzene	3.731	3.743	-0.3	78	0.00
79 T	2-Chlorotoluene	2.063	2.047	0.8	79	0.00
80 T	1,3,5-Trimethylbenzene	2.541	2.597	-2.2	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.171	0.181	-5.8	84	0.00
82 T	4-Chlorotoluene	2.140	2.168	-1.3	80	0.00
83 T	tert-Butylbenzene	2.206	2.166	1.8	75	0.00
84 T	1,2,4-Trimethylbenzene	2.566	2.632	-2.6	79	0.00
85 T	sec-Butylbenzene	3.372	3.394	-0.7	79	0.00
86 T	p-Isopropyltoluene	2.834	2.855	-0.7	78	0.00
87 T	1,3-Dichlorobenzene	1.635	1.629	0.4	78	0.00
88 T	1,4-Dichlorobenzene	1.617	1.581	2.2	77	0.00
89 T	n-Butylbenzene	2.577	2.560	0.7	80	0.00
90 T	Hexachloroethane	0.536	0.524	2.2	88	0.00
91 T	1,2-Dichlorobenzene	1.410	1.392	1.3	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.073	0.079	-8.2	102	0.00
93 T	1,2,4-Trichlorobenzene	0.748	0.748	0.0	90	0.00
94 T	Hexachlorobutadiene	0.354	0.339	4.2	88	0.00
95 T	Naphthalene	1.420	1.474	-3.8	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
Data File : VD075631.D
Acq On : 11 Apr 2023 20:37
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 12 03:00:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 18:54:55 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.662	0.673	-1.7	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6