

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
 Data File : VD075628.D
 Acq On : 11 Apr 2023 19:08
 Operator : KP/SY
 Sample : 02260-09
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1C

Quant Time: Apr 12 02:59:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 11 18:54:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	676	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	2045	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	1255	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.52

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	211	38.048	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	76.100%
35) Dibromofluoromethane	7.805	113	718	59.909	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.820%
50) Toluene-d8	10.269	98	1774	37.737	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	75.480%
62) 4-Bromofluorobenzene	12.581	95	68	3.124	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	6.240%#


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041123\
Data File : VD075628.D
Acq On : 11 Apr 2023 19:08
Operator : KP/SY
Sample : 02260-09
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1C


Quant Time: Apr 12 02:59:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041023S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 11 18:54:55 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VD075628.D
Acq: 11 Apr 2023 19:08

Instrument :
MSVOA_D
ClientSampleId :
SP-1C







Tgt Ion:168 Resp: 676
Ion Ratio Lower Upper
168 100
99 0.0 43.8 65.6#



#33
1,2-Dichloroethane-d4
Concen: 38.048 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD075628.D
Acq: 11 Apr 2023 19:08

Tgt Ion: 65 Resp: 211
Ion Ratio Lower Upper
65 100
67 0.0 0.0 102.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.781 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD075628.D
Acq: 11 Apr 2023 19:08




Instrument :
MSVOA_D
ClientSampleId :
SP-1C


Tgt Ion:114 Resp: 2045
Ion Ratio Lower Upper
114 100
63 0.0 0.0 41.6
88 20.2 0.0 30.6



#35
Dibromofluoromethane
Concen: 59.909 ug/l
RT: 7.805 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VD075628.D
Acq: 11 Apr 2023 19:08

Tgt Ion:113 Resp: 718
Ion Ratio Lower Upper
113 100
111 105.2 81.4 122.0
192 0.0 15.2 22.8#









#50
Toluene-d8
Concen: 37.737 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VD075628.D
Acq: 11 Apr 2023 19:08




Instrument :
MSVOA_D
ClientSampleId :
SP-1C


Tgt Ion: 98 Resp: 1774
Ion Ratio Lower Upper
98 100
100 68.2 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 3.124 ug/l
RT: 12.581 min Scan# 1869
Delta R.T. 0.006 min
Lab File: VD075628.D
Acq: 11 Apr 2023 19:08

Tgt Ion: 95 Resp: 68
Ion Ratio Lower Upper
95 100
174 186.8 0.0 164.0#
176 0.0 0.0 156.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.581 min Scan# 1

Delta R.T. -0.006 min

Lab File: VD075628.D

Acq: 11 Apr 2023 19:08

Instrument :

MSVOA_D

ClientSampleId :

SP-1C





Tgt Ion:117 Resp: 1255

Ion Ratio Lower Upper

117 100

82 45.9 44.0 66.0

119 0.0 25.1 37.7#



#72

1,4-Dichlorobenzene-d4

Concen: 0.000 ug/l

Expected RT: 13.52 min

Lab File: VD075628.D

Acq: 11 Apr 2023 19:08

Tgt Ion: 152

Sig Exp Ratio

152 100

115 60.2

150 182.4

