

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
 Data File : VD075745.D
 Acq On : 19 Apr 2023 13:23
 Operator : KP/SY
 Sample : 02370-01RE
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-04-041423RE

Quant Time: Apr 20 06:00:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	3804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	7947	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	10004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	4286	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	3830	106.073	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	212.140%#
35) Dibromofluoromethane	7.810	113	3470	64.919	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.840%
50) Toluene-d8	10.275	98	8789	41.060	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	82.120%
62) 4-Bromofluorobenzene	12.575	95	2961	52.044	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
Data File : VD075745.D
Acq On : 19 Apr 2023 13:23
Operator : KP/SY
Sample : 02370-01RE
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-04-041423RE

Quant Time: Apr 20 06:00:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

