

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
 Data File : VD075747.D
 Acq On : 19 Apr 2023 14:21
 Operator : KP/SY
 Sample : 02270-15
 Misc : 4.43G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-19-0-2-20230411

Quant Time: Apr 20 06:00:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	214093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	423627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.586	117	368957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	139467	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	109443	53.856	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.720%
35) Dibromofluoromethane	7.810	113	135111	47.419	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.840%
50) Toluene-d8	10.269	98	500605	43.873	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.740%
62) 4-Bromofluorobenzene	12.575	95	117794	38.840	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
Data File : VD075747.D
Acq On : 19 Apr 2023 14:21
Operator : KP/SY
Sample : 02270-15
Misc : 4.43G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-19-0-2-20230411

Quant Time: Apr 20 06:00:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

