

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
 Data File : VD075748.D
 Acq On : 19 Apr 2023 14:50
 Operator : KP/SY
 Sample : 02270-22
 Misc : 4.30G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-15-0-2-20230412

Quant Time: Apr 20 06:00:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	5371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	10627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	13283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	4828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	5517	108.216	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 216.440%#	
35) Dibromofluoromethane	7.816	113	4450	62.258	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.520%	
50) Toluene-d8	10.275	98	12137	42.402	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 84.800%	
62) 4-Bromofluorobenzene	12.575	95	3943	51.827	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 103.660%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
Data File : VD075748.D
Acq On : 19 Apr 2023 14:50
Operator : KP/SY
Sample : 02270-22
Misc : 4.30G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-15-0-2-20230412

Quant Time: Apr 20 06:00:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

