

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
 Data File : VD075753.D
 Acq On : 19 Apr 2023 17:14
 Operator : KP/SY
 Sample : 02404-01RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-2-041823RE

Quant Time: Apr 20 06:01:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	2746	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	5106	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	5538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	1834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	2588	99.291	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 198.580%#	
35) Dibromofluoromethane	7.804	113	2475	72.067	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 144.140%	
50) Toluene-d8	10.269	98	6219	45.219	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 90.440%	
62) 4-Bromofluorobenzene	12.575	95	1736	47.490	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.980%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
Data File : VD075753.D
Acq On : 19 Apr 2023 17:14
Operator : KP/SY
Sample : 02404-01RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-2-041823RE

Quant Time: Apr 20 06:01:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

