

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
 Data File : VD075754.D
 Acq On : 19 Apr 2023 17:43
 Operator : KP/SY
 Sample : 02393-01
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-229

Quant Time: Apr 20 06:01:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	123039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	225044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	191236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	62644	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	66989	57.360	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.720%
35) Dibromofluoromethane	7.805	113	79824	52.736	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.480%
50) Toluene-d8	10.269	98	314001	51.802	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.600%
62) 4-Bromofluorobenzene	12.575	95	55512	34.455	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	68.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041923\
Data File : VD075754.D
Acq On : 19 Apr 2023 17:43
Operator : KP/SY
Sample : 02393-01
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-229

Quant Time: Apr 20 06:01:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

