

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073144.D
 Acq On : 19 May 2022 17:54
 Operator : VA/SY
 Sample : VD0519SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0519SBL01

Quant Time: May 20 01:49:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	287858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	496576	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	519710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	240948	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	166849	52.448	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.900%
35) Dibromofluoromethane	7.908	113	169577	51.752	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.500%
50) Toluene-d8	10.332	98	618338	51.192	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.380%
62) 4-Bromofluorobenzene	12.620	95	232903	54.943	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073144.D
Acq On : 19 May 2022 17:54
Operator : VA/SY
Sample : VD0519SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0519SBL01

Quant Time: May 20 01:49:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

