

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073148.D
 Acq On : 19 May 2022 19:48
 Operator : VA/SY
 Sample : N2889-05RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-211-051-SOIL

Quant Time: May 20 01:50:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	38814	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	63807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	76328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	38706	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	33710	78.587	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.180%
35) Dibromofluoromethane	7.908	113	28927	68.704	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	137.400%
50) Toluene-d8	10.332	98	81448	52.477	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.960%
62) 4-Bromofluorobenzene	12.620	95	33797	62.049	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	124.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073148.D
Acq On : 19 May 2022 19:48
Operator : VA/SY
Sample : N2889-05RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-211-051-SOIL

Quant Time: May 20 01:50:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

