

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073150.D
 Acq On : 19 May 2022 20:46
 Operator : VA/SY
 Sample : N2900-03
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-051722

Quant Time: May 20 01:50:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	254936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	451205	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	486189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	224178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	163006	57.857	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.720%
35) Dibromofluoromethane	7.908	113	158211	53.138	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.280%
50) Toluene-d8	10.332	98	566401	51.607	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.220%
62) 4-Bromofluorobenzene	12.620	95	212303	55.120	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.240%
Target Compounds						
16) Acetone	4.162	43	7950	16.291	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073150.D
Acq On : 19 May 2022 20:46
Operator : VA/SY
Sample : N2900-03
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-051722

Quant Time: May 20 01:50:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

