

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073155.D
 Acq On : 19 May 2022 23:10
 Operator : VA/SY
 Sample : N2842-03
 Misc : 5.80G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P016-SS001-0006-01

Quant Time: May 20 01:51:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	307061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	546850	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	574615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	263370	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	179740	52.967	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.940%
35) Dibromofluoromethane	7.908	113	190469	52.784	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.560%
50) Toluene-d8	10.332	98	682414	51.303	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.600%
62) 4-Bromofluorobenzene	12.620	95	248591	53.253	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073155.D
Acq On : 19 May 2022 23:10
Operator : VA/SY
Sample : N2842-03
Misc : 5.80G/5.00ml/MSVOA_D/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P016-SS001-0006-01

Quant Time: May 20 01:51:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

