

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073156.D
 Acq On : 19 May 2022 23:39
 Operator : VA/SY
 Sample : N2842-04
 Misc : 5.73G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P016-SS001-0612-01

Quant Time: May 20 01:52:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	298152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	536049	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	542590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	246310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	165246	50.150	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.300%
35) Dibromofluoromethane	7.908	113	177206	50.098	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.200%
50) Toluene-d8	10.332	98	656078	50.317	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.640%
62) 4-Bromofluorobenzene	12.620	95	232676	50.848	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.700%
Target Compounds						
2) Dichlorodifluoromethane	1.991	85	3439	1.154	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073156.D
 Acq On : 19 May 2022 23:39
 Operator : VA/SY
 Sample : N2842-04
 Misc : 5.73G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P016-SS001-0612-01

Quant Time: May 20 01:52:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

