

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073157.D
 Acq On : 20 May 2022 00:07
 Operator : VA/SY
 Sample : N2842-05RE
 Misc : 6.42G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P016-SS001-1218-01

Quant Time: May 20 01:52:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	20471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	34702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	41676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	18807	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	19453	85.986	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	171.980%#
35) Dibromofluoromethane	7.908	113	16663	72.768	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	145.540%
50) Toluene-d8	10.332	98	44377	52.573	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.614	95	16705	56.392	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073157.D
Acq On : 20 May 2022 00:07
Operator : VA/SY
Sample : N2842-05RE
Misc : 6.42G/5.00ml/MSVOA_D/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P016-SS001-1218-01

Quant Time: May 20 01:52:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

