

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073175.D
 Acq On : 20 May 2022 10:06
 Operator : VA/SY
 Sample : N2842-16
 Misc : 5.38G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P023-SS001-1824-01

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 05/24/2022
 Supervised By :Mahesh Dadoda 05/24/2022

Quant Time: May 20 13:08:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	322592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	569826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	584188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	272429	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	178350	50.027	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 100.060%			
35) Dibromofluoromethane	7.909	113	189470	50.390	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 100.780%			
50) Toluene-d8	10.332	98	702238	50.664	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 101.320%			
62) 4-Bromofluorobenzene	12.620	95	266792	54.847	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 109.700%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.191	59	11921m	21.552	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073175.D
Acq On : 20 May 2022 10:06
Operator : VA/SY
Sample : N2842-16
Misc : 5.38G/5.00ml/MSVOA_D/SOIL
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P023-SS001-1824-01

Quant Time: May 20 13:08:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/24/2022
Supervised By :Mahesh Dadoda 05/24/2022

