

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073176.D
 Acq On : 20 May 2022 10:34
 Operator : VA/SY
 Sample : N2842-17
 Misc : 3.54G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P023-SS001-2430-01

Quant Time: May 20 13:08:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	159886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	292226	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	318776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	150101	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	124009	70.182	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.360%	
35) Dibromofluoromethane	7.908	113	111840	57.999	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.000%	
50) Toluene-d8	10.332	98	364109	51.224	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.440%	
62) 4-Bromofluorobenzene	12.620	95	137896	55.279	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.560%	
Target Compounds						
11) Tert butyl alcohol	5.173	59	13226	48.245	ug/l	Qvalue # 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073176.D
Acq On : 20 May 2022 10:34
Operator : VA/SY
Sample : N2842-17
Misc : 3.54G/5.00ml/MSVOA_D/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P023-SS001-2430-01

Quant Time: May 20 13:08:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

