

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073178.D
 Acq On : 20 May 2022 11:32
 Operator : VA/SY
 Sample : N2842-19
 Misc : 5.16G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P023-SS003-0006-01

Quant Time: May 20 13:09:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	77949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	150039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	196461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	99253	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	113360	131.593	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 263.180%#	
35) Dibromofluoromethane	7.909	113	75581	76.340	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 152.680%#	
50) Toluene-d8	10.332	98	195283	53.508	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.020%	
62) 4-Bromofluorobenzene	12.620	95	92768	72.430	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 144.860%#	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.173	59	16783	125.571	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073178.D
Acq On : 20 May 2022 11:32
Operator : VA/SY
Sample : N2842-19
Misc : 5.16G/5.00ml/MSVOA_D/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P023-SS003-0006-01

Quant Time: May 20 13:09:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

