

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073143.D
 Acq On : 19 May 2022 16:57
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 20 01:48:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	45.753	8.5	106	0.00
3 P	Chloromethane	50.000	45.794	8.4	106	0.00
4 C	Vinyl Chloride	50.000	46.874	6.3#	103	0.00
5 T	Bromomethane	50.000	53.081	-6.2	106	0.00
6 T	Chloroethane	50.000	49.220	1.6	107	0.00
7 T	Trichlorofluoromethane	50.000	49.219	1.6	107	0.00
8 T	Diethyl Ether	50.000	50.856	-1.7	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.618	2.8	107	0.00
10 T	Methyl Iodide	50.000	50.001	-0.0	102	0.00
11 T	Tert butyl alcohol	250.000	131.624	47.4#	103	0.00
12 CM	1,1-Dichloroethene	50.000	48.979	2.0#	103	0.00
13 T	Acrolein	250.000	232.216	7.1	99	0.00
14 T	Allyl chloride	50.000	49.418	1.2	104	0.00
15 T	Acrylonitrile	250.000	260.423	-4.2	109	0.00
16 T	Acetone	250.000	265.776	-6.3	125	0.00
17 T	Carbon Disulfide	50.000	48.246	3.5	105	0.00
18 T	Methyl Acetate	50.000	49.870	0.3	109	0.00
19 T	Methyl tert-butyl Ether	50.000	52.550	-5.1	108	0.00
20 T	Methylene Chloride	50.000	52.428	-4.9	109	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.716	-1.4	107	0.00
22 T	Diisopropyl ether	50.000	51.479	-3.0	104	0.00
23 T	Vinyl Acetate	250.000	265.818	-6.3	108	0.00
24 P	1,1-Dichloroethane	50.000	48.739	2.5	105	0.00
25 T	2-Butanone	250.000	260.577	-4.2	113	0.00
26 T	2,2-Dichloropropane	50.000	49.140	1.7	106	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.513	1.0	104	0.00
28 T	Bromochloromethane	50.000	48.612	2.8	101	0.00
29 T	Tetrahydrofuran	250.000	261.370	-4.5	109	0.00
30 C	Chloroform	50.000	48.868	2.3#	105	0.00
31 T	Cyclohexane	50.000	48.859	2.3	106	0.00
32 T	1,1,1-Trichloroethane	50.000	48.697	2.6	105	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.935	2.1	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	50.668	-1.3	104	0.00
36 T	1,1-Dichloropropene	50.000	50.978	-2.0	107	0.00
37 T	Ethyl Acetate	50.000	51.913	-3.8	112	0.00
38 T	Carbon Tetrachloride	50.000	50.465	-0.9	104	0.00
39 T	Methylcyclohexane	50.000	53.327	-6.7	108	0.00
40 TM	Benzene	50.000	50.283	-0.6	106	0.00
41 T	Methacrylonitrile	50.000	45.849	8.3	110	0.00
42 TM	1,2-Dichloroethane	50.000	49.728	0.5	105	0.00
43 T	Isopropyl Acetate	50.000	52.585	-5.2	109	0.00
44 TM	Trichloroethene	50.000	51.008	-2.0	107	0.00
45 C	1,2-Dichloropropane	50.000	49.776	0.4#	104	0.00
46 T	Dibromomethane	50.000	49.828	0.3	106	0.00
47 T	Bromodichloromethane	50.000	49.544	0.9	105	0.00
48 T	Methyl methacrylate	50.000	55.993	-12.0	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073143.D
 Acq On : 19 May 2022 16:57
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 20 01:48:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1022.046	-2.2	108	0.00
50 S	Toluene-d8	50.000	51.560	-3.1	104	0.00
51 T	4-Methyl-2-Pentanone	250.000	270.618	-8.2	111	0.00
52 CM	Toluene	50.000	51.722	-3.4#	105	0.00
53 T	t-1,3-Dichloropropene	50.000	51.147	-2.3	108	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.409	-2.8	107	0.00
55 T	1,1,2-Trichloroethane	50.000	50.960	-1.9	108	0.00
56 T	Ethyl methacrylate	50.000	55.613	-11.2	107	0.00
57 T	1,3-Dichloropropane	50.000	51.522	-3.0	106	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	246.809	1.3	109	0.00
59 T	2-Hexanone	250.000	279.921	-12.0	113	0.00
60 T	Dibromochloromethane	50.000	50.855	-1.7	107	0.00
61 T	1,2-Dibromoethane	50.000	51.016	-2.0	108	0.00
62 S	4-Bromofluorobenzene	50.000	51.973	-3.9	107	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	104	0.00
64 T	Tetrachloroethene	50.000	50.418	-0.8	108	0.00
65 PM	Chlorobenzene	50.000	50.205	-0.4	107	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.626	-1.3	106	0.00
67 C	Ethyl Benzene	50.000	52.741	-5.5#	107	0.00
68 T	m/p-Xylenes	100.000	105.792	-5.8	105	0.00
69 T	o-Xylene	50.000	52.997	-6.0	107	0.00
70 T	Styrene	50.000	53.804	-7.6	107	0.00
71 P	Bromoform	50.000	51.711	-3.4	110	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	53.740	-7.5	107	0.00
74 T	N-amyl acetate	50.000	53.428	-6.9	109	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.080	-2.2	109	0.00
76 T	1,2,3-Trichloropropane	50.000	49.284	1.4	104	0.00
77 T	Bromobenzene	50.000	51.448	-2.9	108	0.00
78 T	n-propylbenzene	50.000	53.003	-6.0	106	0.00
79 T	2-Chlorotoluene	50.000	51.929	-3.9	106	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.850	-7.7	108	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	53.225	-6.5	108	0.00
82 T	4-Chlorotoluene	50.000	51.923	-3.8	106	0.00
83 T	tert-Butylbenzene	50.000	54.274	-8.5	109	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.460	-6.9	108	0.00
85 T	sec-Butylbenzene	50.000	53.335	-6.7	108	0.00
86 T	p-Isopropyltoluene	50.000	53.300	-6.6	106	0.00
87 T	1,3-Dichlorobenzene	50.000	51.136	-2.3	107	0.00
88 T	1,4-Dichlorobenzene	50.000	50.017	-0.0	107	0.00
89 T	n-Butylbenzene	50.000	54.022	-8.0	107	0.00
90 T	Hexachloroethane	50.000	50.910	-1.8	107	0.00
91 T	1,2-Dichlorobenzene	50.000	51.669	-3.3	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.827	-3.7	112	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.960	-7.9	109	0.00
94 T	Hexachlorobutadiene	50.000	51.593	-3.2	112	0.00
95 T	Naphthalene	50.000	50.639	-1.3	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073143.D
Acq On : 19 May 2022 16:57
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 20 01:48:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.910	-5.8	107	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6