

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073164.D
 Acq On : 20 May 2022 03:28
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 20 04:27:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.500	0.417	16.6	80	0.00
3 P	Chloromethane	0.565	0.527	6.7	90	0.00
4 C	Vinyl Chloride	0.517	0.464	10.3#	81	0.00
5 T	Bromomethane	0.374	0.348	7.0	87	0.00
6 T	Chloroethane	0.318	0.310	2.5	87	0.00
7 T	Trichlorofluoromethane	0.913	0.859	5.9	84	0.00
8 T	Diethyl Ether	0.251	0.276	-10.0	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.502	7.9	84	0.00
10 T	Methyl Iodide	0.588	0.590	-0.3	85	0.00
11 T	Tert butyl alcohol	0.086	0.043	50.0#	82	0.00
12 CM	1,1-Dichloroethene	0.502	0.485	3.4#	84	0.00
13 T	Acrolein	0.039	0.033	15.4	85	0.00
14 T	Allyl chloride	0.811	0.813	-0.2	87	0.00
15 T	Acrylonitrile	0.115	0.134	-16.5	101	0.00
16 T	Acetone	0.096	0.099	-3.1	100	0.00
17 T	Carbon Disulfide	1.666	1.619	2.8	87	0.00
18 T	Methyl Acetate	0.266	0.292	-9.8	99	0.00
19 T	Methyl tert-butyl Ether	1.145	1.299	-13.4	97	0.00
20 T	Methylene Chloride	0.716	0.702	2.0	100	0.00
21 T	trans-1,2-Dichloroethene	0.608	0.630	-3.6	90	0.00
22 T	Diisopropyl ether	1.700	1.915	-12.6	94	0.00
23 T	Vinyl Acetate	0.907	1.046	-15.3	97	0.00
24 P	1,1-Dichloroethane	1.091	1.143	-4.8	93	0.00
25 T	2-Butanone	0.140	0.160	-14.3	102	0.00
26 T	2,2-Dichloropropane	0.931	0.826	11.3	79	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.731	-8.8	94	0.00
28 T	Bromochloromethane	0.376	0.421	-12.0	96	0.00
29 T	Tetrahydrofuran	0.091	0.105	-15.4	100	0.00
30 C	Chloroform	1.139	1.225	-7.6#	95	0.00
31 T	Cyclohexane	0.966	0.870	9.9	81	0.00
32 T	1,1,1-Trichloroethane	0.997	0.989	0.8	88	0.00
33 S	1,2-Dichloroethane-d4	0.553	0.600	-8.5	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.330	0.347	-5.2	92	0.00
36 T	1,1-Dichloropropene	0.496	0.483	2.6	87	0.00
37 T	Ethyl Acetate	0.196	0.219	-11.7	103	0.00
38 T	Carbon Tetrachloride	0.503	0.485	3.6	85	0.00
39 T	Methylcyclohexane	0.553	0.528	4.5	82	0.00
40 TM	Benzene	1.430	1.487	-4.0	94	0.00
41 T	Methacrylonitrile	0.114	0.129	-13.2	117	0.00
42 TM	1,2-Dichloroethane	0.398	0.432	-8.5	98	0.00
43 T	Isopropyl Acetate	0.362	0.404	-11.6	99	0.00
44 TM	Trichloroethene	0.381	0.388	-1.8	91	0.00
45 C	1,2-Dichloropropane	0.360	0.388	-7.8#	96	0.00
46 T	Dibromomethane	0.198	0.216	-9.1	99	0.00
47 T	Bromodichloromethane	0.506	0.545	-7.7	97	0.00
48 T	Methyl methacrylate	0.182	0.205	-12.6	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
 Data File : VD073164.D
 Acq On : 20 May 2022 03:28
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 20 04:27:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	102	0.01
50 S	Toluene-d8	1.216	1.316	-8.2	93	0.00
51 T	4-Methyl-2-Pentanone	0.188	0.221	-17.6	103	0.00
52 CM	Toluene	0.904	0.974	-7.7#	93	0.00
53 T	t-1,3-Dichloropropene	0.457	0.486	-6.3	96	0.00
54 T	cis-1,3-Dichloropropene	0.547	0.571	-4.4	93	0.00
55 T	1,1,2-Trichloroethane	0.267	0.297	-11.2	100	0.00
56 T	Ethyl methacrylate	0.297	0.353	-18.9	98	0.00
57 T	1,3-Dichloropropane	0.449	0.500	-11.4	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.122	0.138	-13.1	93	0.00
59 T	2-Hexanone	0.124	0.150	-21.0	104	0.00
60 T	Dibromochloromethane	0.334	0.365	-9.3	99	0.00
61 T	1,2-Dibromoethane	0.256	0.285	-11.3	101	0.00
62 S	4-Bromofluorobenzene	0.427	0.459	-7.5	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.327	0.317	3.1	90	0.00
65 PM	Chlorobenzene	1.026	1.074	-4.7	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.402	-5.8	97	0.00
67 C	Ethyl Benzene	1.801	1.876	-4.2#	92	0.00
68 T	m/p-Xylenes	0.700	0.748	-6.9	93	0.00
69 T	o-Xylene	0.643	0.687	-6.8	94	0.00
70 T	Styrene	1.106	1.217	-10.0	95	0.00
71 P	Bromoform	0.196	0.218	-11.2	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.564	3.683	-3.3	91	0.00
74 T	N-amyl acetate	0.778	0.870	-11.8	101	0.00
75 P	1,1,2,2-Tetrachloroethane	0.657	0.720	-9.6	103	0.00
76 T	1,2,3-Trichloropropane	0.491	0.572	-16.5	109	0.00
77 T	Bromobenzene	0.839	0.869	-3.6	96	0.00
78 T	n-propylbenzene	4.475	4.585	-2.5	90	0.00
79 T	2-Chlorotoluene	2.584	2.686	-3.9	94	0.00
80 T	1,3,5-Trimethylbenzene	3.019	3.182	-5.4	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.204	0.216	-5.9	95	0.00
82 T	4-Chlorotoluene	2.708	2.858	-5.5	95	0.00
83 T	tert-Butylbenzene	2.545	2.572	-1.1	90	0.00
84 T	1,2,4-Trimethylbenzene	2.983	3.218	-7.9	96	0.00
85 T	sec-Butylbenzene	3.886	3.913	-0.7	90	0.00
86 T	p-Isopropyltoluene	3.167	3.257	-2.8	91	0.00
87 T	1,3-Dichlorobenzene	1.691	1.733	-2.5	95	0.00
88 T	1,4-Dichlorobenzene	1.707	1.727	-1.2	96	0.00
89 T	n-Butylbenzene	3.018	3.019	-0.0	88	0.00
90 T	Hexachloroethane	0.655	0.655	0.0	93	0.00
91 T	1,2-Dichlorobenzene	1.463	1.526	-4.3	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.106	-5.0	101	0.00
93 T	1,2,4-Trichlorobenzene	0.831	0.867	-4.3	93	0.00
94 T	Hexachlorobutadiene	0.468	0.450	3.8	92	0.00
95 T	Naphthalene	1.466	1.692	-15.4	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051922\
Data File : VD073164.D
Acq On : 20 May 2022 03:28
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 20 04:27:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.737	0.787	-6.8	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6