

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062122\
 Data File : VD073644.D
 Acq On : 21 Jun 2022 13:55
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jun 22 06:01:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 05:24:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	266140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	471744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	434109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	189727	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	151456	49.459	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.920%
35) Dibromofluoromethane	7.908	113	155518	45.443	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.880%
50) Toluene-d8	10.332	98	548509	44.780	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.560%
62) 4-Bromofluorobenzene	12.620	95	182300	40.584	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062122\
Data File : VD073644.D
Acq On : 21 Jun 2022 13:55
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jun 22 06:01:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 05:24:45 2022
Response via : Initial Calibration

