

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062122\
 Data File : VD073646.D
 Acq On : 21 Jun 2022 15:23
 Operator : VA/SY
 Sample : N3428-01
 Misc : 5.21G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-01-062122

Quant Time: Jun 22 06:01:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 05:24:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	253103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	456655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	421971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	189561	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	146026	50.142	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.280%
35) Dibromofluoromethane	7.902	113	153381	46.300	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.600%
50) Toluene-d8	10.332	98	546899	46.124	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.240%
62) 4-Bromofluorobenzene	12.620	95	175186	40.288	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.580%
Target Compounds						Qvalue
20) Methylene Chloride	4.955	84	36544	7.459	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062122\
Data File : VD073646.D
Acq On : 21 Jun 2022 15:23
Operator : VA/SY
Sample : N3428-01
Misc : 5.21G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-01-062122

Quant Time: Jun 22 06:01:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062022S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 05:24:45 2022
Response via : Initial Calibration

