

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
 Data File : VD079292.D
 Acq On : 01 Jul 2024 13:28
 Operator : RP/MD
 Sample : P3127-07
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-196-GW-770-772

Quant Time: Jul 01 15:36:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 29 06:28:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	95170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	176825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	149580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	41140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	54253	59.033	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.060%
35) Dibromofluoromethane	7.798	113	63051	54.323	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.640%
50) Toluene-d8	10.269	98	185673	44.467	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.940%
62) 4-Bromofluorobenzene	12.575	95	51892	38.756	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.520%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.793	84	11739	2.258	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
Data File : VD079292.D
Acq On : 01 Jul 2024 13:28
Operator : RP/MD
Sample : P3127-07
Misc : 5.04G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-196-GW-770-772

Quant Time: Jul 01 15:36:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 29 06:28:31 2024
Response via : Initial Calibration

