

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
 Data File : VD079296.D
 Acq On : 01 Jul 2024 15:18
 Operator : RP/MD
 Sample : P3135-02
 Misc : 3.45G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-628-VOC-01

Quant Time: Jul 02 05:58:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 29 06:28:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	139649	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	300273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	286161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	114232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	87283	65.153	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.300%
35) Dibromofluoromethane	7.805	113	97096	49.263	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.520%
50) Toluene-d8	10.269	98	293121	41.339	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	82.680%
62) 4-Bromofluorobenzene	12.569	95	102903	45.257	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.520%
Target Compounds						
20) Methylene Chloride	4.793	84	16606	1.952	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
Data File : VD079296.D
Acq On : 01 Jul 2024 15:18
Operator : RP/MD
Sample : P3135-02
Misc : 3.45G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-628-VOC-01

Quant Time: Jul 02 05:58:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 29 06:28:31 2024
Response via : Initial Calibration

