

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
 Data File : VD079298.D
 Acq On : 01 Jul 2024 16:13
 Operator : RP/MD
 Sample : P3077-01
 Misc : 8.20G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 1-N-1

Quant Time: Jul 02 05:58:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 29 06:28:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	119484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	256744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	222488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	76576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	75156	65.598	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.200%
35) Dibromofluoromethane	7.804	113	87944	52.184	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.360%
50) Toluene-d8	10.263	98	217815	35.927	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	71.860%
62) 4-Bromofluorobenzene	12.569	95	81361	41.850	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.700%
Target Compounds						
64) Tetrachloroethene	10.810	164	8378	6.044	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
Data File : VD079298.D
Acq On : 01 Jul 2024 16:13
Operator : RP/MD
Sample : P3077-01
Misc : 8.20G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
1-N-1

Quant Time: Jul 02 05:58:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 29 06:28:31 2024
Response via : Initial Calibration

