

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
 Data File : VD079300.D
 Acq On : 01 Jul 2024 17:08
 Operator : RP/MD
 Sample : P3135-08
 Misc : 4.49G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-628-VOC-02

Quant Time: Jul 02 05:59:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 29 06:28:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	117399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	243137	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	216422	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	77767	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	75786	67.439	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.880%
35) Dibromofluoromethane	7.804	113	85873	53.807	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.620%
50) Toluene-d8	10.263	98	240420	41.875	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	83.740%
62) 4-Bromofluorobenzene	12.569	95	77761	42.237	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.480%
Target Compounds						
					Qvalue	
16) Acetone	4.028	43	3654	13.114	ug/l #	81
20) Methylene Chloride	4.793	84	14105	2.037	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
Data File : VD079300.D
Acq On : 01 Jul 2024 17:08
Operator : RP/MD
Sample : P3135-08
Misc : 4.49G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-628-VOC-02

Quant Time: Jul 02 05:59:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 29 06:28:31 2024
Response via : Initial Calibration

