

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
 Data File : VD079301.D
 Acq On : 01 Jul 2024 17:35
 Operator : RP/MD
 Sample : P3111-02
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: Jul 02 06:00:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 29 06:28:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	107384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	212991	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	172772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	47255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	81066	79.621	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	159.240%
35) Dibromofluoromethane	7.804	113	92858	66.419	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	132.840%
50) Toluene-d8	10.269	98	242479	48.211	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.420%
62) 4-Bromofluorobenzene	12.569	95	64146	39.773	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.540%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.805	84	18070	5.350	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
Data File : VD079301.D
Acq On : 01 Jul 2024 17:35
Operator : RP/MD
Sample : P3111-02
Misc : 5.02G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: Jul 02 06:00:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 29 06:28:31 2024
Response via : Initial Calibration

