

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
 Data File : VD079302.D
 Acq On : 01 Jul 2024 18:03
 Operator : RP/MD
 Sample : P3112-02
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: Jul 02 06:00:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 29 06:28:31 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	19296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	41500	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	38968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	11598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	24198	135.213	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 270.420%#	
35) Dibromofluoromethane	7.805	113	21764	79.896	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 159.800%#	
50) Toluene-d8	10.269	98	41583	42.433	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 84.860%	
62) 4-Bromofluorobenzene	12.569	95	14178	45.118	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 90.240%	
Target Compounds						Qvalue
20) Methylene Chloride	4.793	84	4608	10.206	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070124\
Data File : VD079302.D
Acq On : 01 Jul 2024 18:03
Operator : RP/MD
Sample : P3112-02
Misc : 5.02G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: Jul 02 06:00:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062824S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 29 06:28:31 2024
Response via : Initial Calibration

