


Dichlorodifluoromethane

Response Ratio



$$\text{Response} = 4.269\text{e-}001 * \text{Amt} + 3.451\text{e-}002$$


Coef of Det (r^2) = 0.997376 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M

Calibration Table Last Updated: Sat Aug 06 00:36:42 2022

2-Chloroethyl Vinyl ether

Response Ratio



$$\text{Response} = 1.269\text{e-}001 * \text{Amt} - 4.327\text{e-}002$$


Coef of Det (r^2) = 0.999286 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M

Calibration Table Last Updated: Sat Aug 06 00:36:42 2022

Toluene-d8

Response Ratio



$$\text{Response} = 7.740\text{e-}001 * \text{Amt} + 7.372\text{e-}002$$


Coef of Det (r^2) = 0.996329 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M

Calibration Table Last Updated: Sat Aug 06 00:36:42 2022

1,2-Dichloroethane-d4

Response Ratio



$$\text{Response} = 3.296\text{e-}001 * \text{Amt} + 5.592\text{e-}002$$


Coef of Det (r^2) = 0.996301 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M

Calibration Table Last Updated: Sat Aug 06 00:36:42 2022

Methylene Chloride

Response Ratio




$$\text{Response} = 4.740\text{e-}001 * \text{Amt} + 1.270\text{e-}001$$

Coef of Det (r^2) = 0.998033 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M

Calibration Table Last Updated: Sat Aug 06 00:36:42 2022


Methyl Acetate



Response = 2.039e-001 * Amt + 1.557e-002
Coef of Det (r^2) = 0.999202 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M
Calibration Table Last Updated: Sat Aug 06 00:36:42 2022

Acetone

Response Ratio




$$\text{Response} = 6.498\text{e-}002 * \text{Amt} + 3.980\text{e-}002$$

Coef of Det (r^2) = 0.994157 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M

Calibration Table Last Updated: Sat Aug 06 00:36:42 2022


Acrolein



$\text{Response} = 3.775\text{e-}002 * \text{Amt} - 2.390\text{e-}002$
 Coef of Det (r^2) = 0.993115 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M
 Calibration Table Last Updated: Sat Aug 06 00:36:42 2022

Tert butyl alcohol

Response Ratio



$$\text{Response} = 3.115\text{e-}002 * \text{Amt} + 1.375\text{e-}002$$

Coef of Det (r^2) = 0.995284 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D080522S.M

Calibration Table Last Updated: Sat Aug 06 00:36:42 2022