

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080822\
 Data File : VD073977.D
 Acq On : 08 Aug 2022 17:23
 Operator : VA/SY
 Sample : N3995-03
 Misc : 4.16G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-3

Quant Time: Aug 08 18:30:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 06 00:36:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	58403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	143895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	156293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	60967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	57262	140.248	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 280.500%#		
35) Dibromofluoromethane	7.908	113	55350	62.436	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 124.880%		
50) Toluene-d8	10.332	98	120112	49.161	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 98.320%		
62) 4-Bromofluorobenzene	12.620	95	52534	44.465	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 88.940%		
Target Compounds						
						Qvalue
16) Acetone	4.156	43	6165	50.606	ug/l	98
20) Methylene Chloride	4.956	84	19697	22.182	ug/l	93
81) trans-1,4-Dichloro-2-b...	12.614	75	28486	127.298	ug/l #	11

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080822\
Data File : VD073977.D
Acq On : 08 Aug 2022 17:23
Operator : VA/SY
Sample : N3995-03
Misc : 4.16G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-3

Quant Time: Aug 08 18:30:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080522S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 06 00:36:42 2022
Response via : Initial Calibration

