

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080822\
 Data File : VD073980.D
 Acq On : 08 Aug 2022 18:53
 Operator : VA/SY
 Sample : N4023-04RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP2-1-6RE

Quant Time: Aug 08 19:13:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 06 00:36:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	10577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	18139	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	18718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	8800	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	8841	118.314	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	236.620%#
35) Dibromofluoromethane	7.903	113	7248	64.858	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.720%
50) Toluene-d8	10.326	98	14834	48.068	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.140%
62) 4-Bromofluorobenzene	12.620	95	7303	49.036	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080822\
Data File : VD073980.D
Acq On : 08 Aug 2022 18:53
Operator : VA/SY
Sample : N4023-04RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP2-1-6RE

Quant Time: Aug 08 19:13:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080522S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 06 00:36:42 2022
Response via : Initial Calibration

