

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080822\
 Data File : VD073983.D
 Acq On : 08 Aug 2022 20:23
 Operator : VA/SY
 Sample : N3995-02
 Misc : 4.27G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-2

Quant Time: Aug 08 23:27:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 06 00:36:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	98389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	172517	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	145608	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	54952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	50247	68.987	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 137.980%			
35) Dibromofluoromethane	7.902	113	58631	55.164	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 110.320%			
50) Toluene-d8	10.332	98	131145	44.346	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 88.700%			
62) 4-Bromofluorobenzene	12.620	95	54399	38.405	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 76.800%			
Target Compounds						
20) Methylene Chloride	4.956	84	23737	12.054	ug/l	94
95) Naphthalene	15.373	128	14271	8.185	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080822\
Data File : VD073983.D
Acq On : 08 Aug 2022 20:23
Operator : VA/SY
Sample : N3995-02
Misc : 4.27G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-2

Quant Time: Aug 08 23:27:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080522S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 06 00:36:42 2022
Response via : Initial Calibration

