



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080822\
 Data File : VD073972.D
 Acq On : 08 Aug 2022 10:04
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080522S.M
 Title : SW846 8260
 Last Update : Sat Aug 06 00:36:42 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	12700	PASS
75	95	30	60	52.9	35661	PASS
95	95	100	100	100.0	67456	PASS
96	95	5	9	6.7	4522	PASS
173	174	0.00	2	1.9	1076	PASS
174	95	50	100	85.8	57899	PASS
175	174	5	9	7.8	4519	PASS
176	174	95	101	96.6	55915	PASS
177	176	5	9	6.8	3829	PASS