

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091222\
 Data File : VD074259.D
 Acq On : 12 Sep 2022 18:46
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 13 04:43:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.434	0.379	12.7	88	0.00
3 P	Chloromethane	0.374	0.346	7.5	92	0.00
4 C	Vinyl Chloride	0.318	0.326	-2.5#	94	0.00
5 T	Bromomethane	0.168	0.147	12.5	82	0.00
6 T	Chloroethane	0.190	0.211	-11.1	108	0.00
7 T	Trichlorofluoromethane	0.886	0.647	27.0#	68	0.00
8 T	Diethyl Ether	0.286	0.299	-4.5	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.570	0.554	2.8	90	0.00
10 T	Methyl Iodide	0.679	0.618	9.0	79	0.00
11 T	Tert butyl alcohol	0.042	0.042	0.0	96	0.00
12 CM	1,1-Dichloroethene	0.550	0.537	2.4#	90	0.00
13 T	Acrolein	0.044	0.040	9.1	79	0.00
14 T	Allyl chloride	0.854	0.812	4.9	87	0.00
15 T	Acrylonitrile	0.133	0.144	-8.3	94	0.00
16 T	Acetone	0.105	0.113	-7.6	83	0.00
17 T	Carbon Disulfide	1.587	1.491	6.0	88	-0.01
18 T	Methyl Acetate	0.352	0.350	0.6	94	0.00
19 T	Methyl tert-butyl Ether	1.333	1.409	-5.7	94	0.00
20 T	Methylene Chloride	0.814	0.778	4.4	107	0.00
21 T	trans-1,2-Dichloroethene	0.621	0.620	0.2	91	-0.01
22 T	Diisopropyl ether	1.812	1.908	-5.3	93	0.00
23 T	Vinyl Acetate	0.856	0.989	-15.5	97	0.00
24 P	1,1-Dichloroethane	1.119	1.153	-3.0	92	0.00
25 T	2-Butanone	0.175	0.181	-3.4	90	0.00
26 T	2,2-Dichloropropane	0.978	0.941	3.8	87	0.00
27 T	cis-1,2-Dichloroethene	0.709	0.726	-2.4	93	0.00
28 T	Bromochloromethane	0.344	0.371	-7.8	88	0.00
29 T	Tetrahydrofuran	0.108	0.119	-10.2	97	0.00
30 C	Chloroform	1.111	1.170	-5.3#	96	0.00
31 T	Cyclohexane	1.053	0.969	8.0	89	0.00
32 T	1,1,1-Trichloroethane	0.975	0.999	-2.5	92	0.00
33 S	1,2-Dichloroethane-d4	0.510	0.524	-2.7	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.364	0.338	7.1	89	0.00
36 T	1,1-Dichloropropene	0.541	0.538	0.6	92	0.00
37 T	Ethyl Acetate	0.235	0.257	-9.4	99	0.00
38 T	Carbon Tetrachloride	0.490	0.518	-5.7	94	0.00
39 T	Methylcyclohexane	0.670	0.652	2.7	90	0.00
40 TM	Benzene	1.498	1.526	-1.9	92	0.00
41 T	Methacrylonitrile	0.131	0.148	-13.0	90	0.00
42 TM	1,2-Dichloroethane	0.396	0.418	-5.6	96	0.00
43 T	Isopropyl Acetate	0.453	0.483	-6.6	95	0.00
44 TM	Trichloroethene	0.416	0.417	-0.2	91	0.00
45 C	1,2-Dichloropropane	0.389	0.403	-3.6#	94	0.00
46 T	Dibromomethane	0.202	0.209	-3.5	93	0.00
47 T	Bromodichloromethane	0.508	0.544	-7.1	96	0.00
48 T	Methyl methacrylate	0.214	0.231	-7.9	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091222\
 Data File : VD074259.D
 Acq On : 12 Sep 2022 18:46
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 13 04:43:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	99	0.00
50 S	Toluene-d8	1.226	1.290	-5.2	87	0.00
51 T	4-Methyl-2-Pentanone	0.229	0.255	-11.4	97	0.00
52 CM	Toluene	0.961	0.980	-2.0#	92	0.00
53 T	t-1,3-Dichloropropene	0.479	0.521	-8.8	95	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.619	-4.7	94	0.00
55 T	1,1,2-Trichloroethane	0.279	0.300	-7.5	94	0.00
56 T	Ethyl methacrylate	0.371	0.410	-10.5	95	0.00
57 T	1,3-Dichloropropane	0.493	0.522	-5.9	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.077	0.065	15.6	96	0.00
59 T	2-Hexanone	0.156	0.175	-12.2	95	0.00
60 T	Dibromochloromethane	0.340	0.368	-8.2	96	0.00
61 T	1,2-Dibromoethane	0.266	0.287	-7.9	95	0.00
62 S	4-Bromofluorobenzene	0.507	0.411	18.9	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.382	0.369	3.4	89	0.00
65 PM	Chlorobenzene	1.124	1.162	-3.4	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.403	0.426	-5.7	94	0.00
67 C	Ethyl Benzene	2.100	2.170	-3.3#	93	0.00
68 T	m/p-Xylenes	0.797	0.827	-3.8	93	0.00
69 T	o-Xylene	0.753	0.780	-3.6	92	0.00
70 T	Styrene	1.256	1.341	-6.8	94	0.00
71 P	Bromoform	0.224	0.249	-11.2	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.230	4.236	-0.1	93	0.00
74 T	N-amyl acetate	0.996	1.068	-7.2	96	0.00
75 P	1,1,2,2-Tetrachloroethane	0.770	0.807	-4.8	96	0.00
76 T	1,2,3-Trichloropropane	0.513	0.510	0.6	84	0.00
77 T	Bromobenzene	0.928	0.961	-3.6	97	0.00
78 T	n-propylbenzene	5.187	5.198	-0.2	94	0.00
79 T	2-Chlorotoluene	2.806	2.797	0.3	93	0.00
80 T	1,3,5-Trimethylbenzene	3.505	3.459	1.3	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.208	0.222	-6.7	97	0.00
82 T	4-Chlorotoluene	2.857	2.845	0.4	93	0.00
83 T	tert-Butylbenzene	3.068	3.035	1.1	93	0.00
84 T	1,2,4-Trimethylbenzene	3.403	3.433	-0.9	94	0.00
85 T	sec-Butylbenzene	4.647	4.609	0.8	93	0.00
86 T	p-Isopropyltoluene	3.796	3.871	-2.0	95	0.00
87 T	1,3-Dichlorobenzene	1.893	1.909	-0.8	95	0.00
88 T	1,4-Dichlorobenzene	1.862	1.879	-0.9	95	0.00
89 T	n-Butylbenzene	3.709	3.784	-2.0	95	0.00
90 T	Hexachloroethane	0.694	0.718	-3.5	96	0.00
91 T	1,2-Dichlorobenzene	1.647	1.665	-1.1	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.108	0.119	-10.2	98	0.00
93 T	1,2,4-Trichlorobenzene	1.138	1.155	-1.5	97	0.00
94 T	Hexachlorobutadiene	0.650	0.659	-1.4	95	0.00
95 T	Naphthalene	2.127	2.302	-8.2	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091222\
Data File : VD074259.D
Acq On : 12 Sep 2022 18:46
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Sep 13 04:43:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.977	1.024	-4.8	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6