

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
 Data File : VD074347.D
 Acq On : 19 Sep 2022 14:22
 Operator : VA/SY
 Sample : N4584-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 33811

Quant Time: Sep 19 23:01:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	103027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	178149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	156344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	68085	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	52731	52.547	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.100%	
35) Dibromofluoromethane	7.811	113	50752	43.256	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 86.520%	
50) Toluene-d8	10.269	98	202701	50.069	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.140%	
62) 4-Bromofluorobenzene	12.569	95	56201	39.796	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.600%	
Target Compounds						
20) Methylene Chloride	4.799	84	28576	13.888	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
Data File : VD074347.D
Acq On : 19 Sep 2022 14:22
Operator : VA/SY
Sample : N4584-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
33811

Quant Time: Sep 19 23:01:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

