

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
 Data File : VD074355.D
 Acq On : 19 Sep 2022 19:26
 Operator : VA/SY
 Sample : N4656-21
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-3

Quant Time: Sep 19 23:02:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	104605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	182395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	153003	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	56413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	43760	42.949	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.900%
35) Dibromofluoromethane	7.805	113	49081	40.858	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.720%
50) Toluene-d8	10.269	98	189371	45.687	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.380%
62) 4-Bromofluorobenzene	12.569	95	47532	31.751	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	63.500%
Target Compounds						
20) Methylene Chloride	4.793	84	29272	14.107	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
Data File : VD074355.D
Acq On : 19 Sep 2022 19:26
Operator : VA/SY
Sample : N4656-21
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-3

Quant Time: Sep 19 23:02:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

