

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
 Data File : VD074356.D
 Acq On : 19 Sep 2022 19:55
 Operator : VA/SY
 Sample : N4656-31
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-4

Quant Time: Sep 19 23:03:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	103047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	182310	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	161174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	75508	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	44645	44.481	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.960%
35) Dibromofluoromethane	7.804	113	49215	40.988	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.980%
50) Toluene-d8	10.269	98	196111	47.335	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.569	95	55965	38.550	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.100%
Target Compounds						
20) Methylene Chloride	4.793	84	29249	14.461	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091922\
Data File : VD074356.D
Acq On : 19 Sep 2022 19:55
Operator : VA/SY
Sample : N4656-31
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-4

Quant Time: Sep 19 23:03:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

